

BASIC RADIATION AIR MONITOR

Survey Meter • Wipe Counter • also Chem or Bio Aerosols

Model ~ AIR-TBM

FEATURES:

- ALPHA, BETA-GAMMA
- IDEAL FOR FIRST RESPONDERS
- REAL TIME
- PORTABLE
- ELECTRONICS – TBM-3SR-D
- EASY CHANGE FILTER
- AC PUMP AP-5020E
- **OPTIONAL:** BATTERY POWERED PUMP
- **OPTIONAL:** ALL AC OPERATION OR ALL BATTERY OPERATION



TBM-3SR-D

APPLICATION:

• QUADRUPLE USE:

- AIR MONITOR- RADIATION
- WIPE COUNTER
- SURFACE MONITOR
- PERSONAL AIR SAMPLER



TBM-3SR-D WITH BATTERY PUMP



AC PUMP, HOSE & FILTER HOLDER

DESCRIPTION:

The AIR-TBM uses standard 57-70 mm filters in T/A's easy change, no leak filter, to trap any airborne radioactivity. The filter is under constant surveillance via pancake GM tube.

Air is drawn through the system by a continuous duty pump.

The unit is completely self-contained and portable. Room air is drawn in through the charcoal impregnated filter paper. Then through the pump and its 3 ft. exhaust hose that can be extended.

The AIR-TBM air sample filter holder from the AC pump can be connected to the TBM-3SR-D through a unique "groove and track" bracket on bottom of TBM-3SR-D. This provides attachment of the air sampler filter holder for read-out with the TBM-3SR-D. The TBM-3SR-D is itself a compact surface monitor and survey meter.



TECHNICAL ASSOCIATES

Divisions of



US NUCLEAR CORP

OVERHOFF TECHNOLOGY

7051 Eton Ave., Canoga Park, CA 91303
818-883-7043 (Phone) 818-883-6103 (Fax)

tagold@nwc.net

WWW.TECH-ASSOCIATES.COM

BASIC RADIATION AIR MONITOR

Survey Meter • Wipe Counter • also Chem or Bio Aerosols

Model ~ AIR-TBM

SPECIFICATIONS:

- Measures:** Airborne radioactivity including aerosol, particulate and gases excluding Noble Gases.
- Sensitivity:** Limit of sensitivity (99% confidence level) 1 nanoCi collected on filter paper.

TIME	SENSITIVITY
1 Hour	3×10^{-9} μ Ci/cc
24 Hours	1×10^{-10} μ Ci/cc
5 Days	2×10^{-11} μ Ci/cc

- 2×10^{-11} μ Ci/cc after 5 days with pump running or 1×10^{-10} μ Ci/cc after 24 hours
- 3×10^{-9} μ Ci/cc after pump runs for 1 hour.
- This level is below 2/10th of the NRC Title 10 limit for 40 Hr. exposure to most hazardous Beta and/or Gamma emitter.

- Detector:** 2" diameter. Thin window 1-3/4" x 1.5-2 mg/cm² pancake T-1190 GM detector
- Background:** Typically 25 cpm
- Flow Rate:** Calibrated at factory. See chart below. Positive displacement pump.
- Filters:** Pack of 50 filters, charcoal impregnated
Optional: Can use membrane, paper or glass fiber filters.
- Case:** Electronics housed in instrument-case, carrying handle included.
- Electronics:** **TBM-3SR-D** Digital Ratemeter.
- Ranges:** 0.01 to 100 mR/hr.
- Meter Response Time:** 2 seconds
- Pump Power:** AC. 115V, 50-60 Hz (230 V optional)
Optional: 6 AA batteries

Pump	Model #	Flow Rate	Weight
Battery Operated	MINI-AIR-15	4ft ³ /hour (2 lpm)	1 lb
AC Operated	AP-5020E	0.5ft ³ /minute (15 lpm)	6 lb

WEIGHT & DIMENSIONS:

- Dimensions TBM-3SR-D:** 3" W x 5" H x 6"D
- Weight TBM-3SR-D:** Electronics – 1.5 lbs
- Dimensions AC Pump:** 5" W x 7" H x 10"L (Fold Down Handle)
- Weight AC Pump:** 6 lbs

OPTIONS:

Inlet hose to allow monitoring of closed spaces.



TECHNICAL ASSOCIATES

Divisions of



US NUCLEAR CORP

OVERHOFF TECHNOLOGY

7051 Eton Ave., Canoga Park, CA 91303
818-883-7043 (Phone) 818-883-6103 (Fax)

tagold@nwc.net

WWW.TECH-ASSOCIATES.COM