



# Hand, Shoe and Clothing Scintillation Monitor

**Model: HSM-10ABS**

## Application

Models **HSM-10ABS** Hand and Shoe Monitor is a health physics instruments.

Designed to provide completely automatic detection and measurement of Beta-Gamma plus Alpha contamination on the palms and backs of each hand, the bottom surfaces of both shoes and optional: the hair and clothing.

Extremely simple operation allows the individual worker to perform this check without direct supervision of the health physicist.

## Features

- Dual Scintillation Detectors Separates and Simultaneously Detects Beta, Low Energy Beta and Alpha
- (3) Separate Read Outs
- Full Dynamic Background Subtract
- Color Monitor Read-Out
- Alarm Settings as Low as Background
- Audio / Visual Alarms
- Incomplete Check Warning
- External Clothing Probe with Separate Circuit, Meter, and Speaker
- Modular Electronics
- Built-In Computer
- Data Archive and Retrieval
- USB / Ethernet Ports
- IP41

## Description

The **HSM-10 Series** monitors feature large full color LCD display screen that enable the user to tell - in one look - whether he is contaminated, and where.

Numerical read-outs are provided, showing the separate counts on each hand and the separate count for each shoe.

An outstanding feature of the instrument is the low warning levels, permitting alarm settings as low as the background count.

A separate rate meter circuit is provided for the EXTERNAL CLOTHING PROBE, including a speaker for audible check and a meter for visual check.

All the detecting elements and sections exposed to contamination are designed for easy decontamination. A roll of kraft paper is mounted in the base for use over the shoe deck preventing



**TECHNICAL ASSOCIATES  
OVERHOFF TECHNOLOGY**

7051 ETON AVENUE, CANOGA PARK, CALIFORNIA 91303

PHONE: 818-883-7043 | FAX: 818-883-6103

SALES@USNUCLEARCORP.COM | TECH-ASSOCIATES.COM | USNUCLEARCORP.COM

DIVISIONS OF



USNUCLEARCORP

OTCQB-UCLC

# Hand, Shoe and Clothing Scintillation Monitor

## Model: HSM-10ABS

### Operations

No training or particular skill is required in the operation of **HSM-10 Series** monitors. The user inserts hands into the two waist-high probe openings.

The counting process is started and continued by pressure of the fingertips at the rear of the probe openings, and by the weight of the person on the shoe deck. As soon as the counting starts, the "Ready for Use" message is replaced by "Counter in Operation".

**Note:** If at any time during the counting cycle, the user removes either hand from the hand probe or steps off the shoe deck, the cycle will automatically stop and the orange message,

"Check Incomplete - Reset and Repeat" will light up. This positive warning prevents erroneous readings.

As soon as the preset time has elapsed, the "Counter in Operation" is replaced by either the green "Check O.K." panel or the bold red "Decontamination Required". Message is dependent upon whether the radioactivity present is above or below the alarm setting.

If decontamination is required, bold red panels marked "Left Hand", "Right Hand", or "Left or Right Shoes" will light up, indicating the location of contamination.

The degree of contamination is indicated on the digital readout. After the check is completed, the instrument is automatically ready for the next user.

**\*\*The CLOTHING PROBE**, is used to detect contamination on hair and clothing. The output of the clothing probe is read visually on its own count rate meter and audibly by the speaker.

The clothing Probe consists of the extremely sensitive T/A **Model P-15 Pancake Unit**. It has a wide face and a thin window and high sensitivity for fast checking of surfaces.

### Specifications: HSM-10ABS – Beta and Gamma

<b>Sensitivity:</b>	Less than .01 $\mu\text{C}$ of 0.2 MeV or higher beta.
<b>Detectors:</b>	Dual phosphor ZnS Alpha and plastic Beta/Gamma scintillation detectors. Two (2) per hand and two (2) per foot totaling eight (8). <b>Optional</b> Clothing probe is a T/A Model P-15 with a T-1190 pancake Alpha, Beta, Gamma sensitive GM tube.
<b>Counters:</b>	Three separate five-digit on-screen Counters for hands (2) and shoe (2), clothing (1)
<b>Shielding:</b>	Hand and Shoe detectors are shielded by 1/4" lead.
<b>Count Time:</b>	Resettable 10, 20 or 30 seconds.
<b>Count Range:</b>	0-99,999 counts Digital each hand or both feet; Clothing probe has four digital output: 0-500,000 cpm.
<b>Alarm:</b>	Audio / Visual Continuously settable to 10,000 total counts, either hand or either shoe. Loss of signal
<b>Activation Switch:</b>	Correct hand position
<b>Readout:</b>	User Settable Units: cps, sv, cpm, Bq, $\mu\text{Ci}$ , etc. Ready for Use, Counter in Operation, Check Incomplete, Reset and Repeat, Check O.K., Decontamination Required, Left Hand, Right Hand, Right Shoe, Left Shoe. A 5" meter for clothing probe, plus audible signal.
<b>Environment:</b>	Temperature: <45° C; <b>Optional:</b> 0°-55° C; RH: <90%



## TECHNICAL ASSOCIATES OVERHOFF TECHNOLOGY

7051 ETON AVENUE, CANOGA PARK, CALIFORNIA 91303

PHONE: 818-883-7043 | FAX: 818-883-6103

SALES@USNUCLEARCORP.COM | TECH-ASSOCIATES.COM | USNUCLEARCORP.COM

DIVISIONS OF



USNUCLEARCORP

OTCQB-UCLE

# Hand, Shoe and Clothing Scintillation Monitor

**Model: HSM-10ABS**

<b>HSM SERIES ~ HAND, SHOE &amp; CLOTHING MONITOR COMPARISON</b>			
<b>MODEL</b>	<b>DETECTION</b>	<b>DETECTOR TYPE</b>	<b>EXTERNAL CLOTHING PROBE</b>
<b>HSM-10A</b>	BETA / GAMMA	THIN SIDE WALL GM (4) GM T-1100 Per hand; (2) GM T-1200 Per foot  TOTAL (8) OPTIONAL (1) each additional GM T-1200 for each foot TOTAL (12)	YES TA – P15; GM T-1190
<b>HSM-10B</b>	ALPHA, BETA, GAMMA	THIN WINDOW PANCAKE GM T-1190 (4) Per hand (2) Per foot TOTAL (12)	YES TA – P15; GM T-1190
<b>HSM-10ABS</b>	SEPARATE & SIMULTANEOUS ALPHA & BETA	SCINTILLATION Dual phosphor ZnS Alpha and plastic Beta/Gamma scintillation detectors. (2) Per hand; (2) Per foot. TOTAL (8)	YES TA – P15; GM T-1190
<b>HSM-10G</b>	SEPARATE & SIMULTANEOUS ALPHA & BETA	GAS FLOW PROPORTIONAL 10-3/4" x 6-1/4" each side – (2) Per hand detectors. 16-1/4" x 6-1/4" each shoe – (1) Per shoe. TOTAL (6)	NO
<b>HSM-10AS (HAND &amp; CLOTHING ONLY)</b>	ALPHA	ALPHA SCINTILLATION ZnS(Ag (2) Per hand. TOTAL (4)	YES TA PAS-8 – ZnS(Ag)
<b>HSM-P-20 (Portable HAND, FOOT &amp; CLOTHING)</b>	BETA / GAMMA OPTIONAL - ALPHA	THIN WINDOW PANCAKE GM T-1190 (2) Portable Hand / Clothing Detector (2) Portable Shoe Detector. TOTAL (4)	YES TA P-15 GM T-1190



## TECHNICAL ASSOCIATES OVERHOFF TECHNOLOGY

7051 ETON AVENUE, CANOGA PARK, CALIFORNIA 91303  
PHONE: 818-883-7043 | FAX: 818-883-6103

SALES@USNUCLEARCORP.COM | TECH-ASSOCIATES.COM | USNUCLEARCORP.COM

DIVISIONS OF



USNUCLEARCORP

OTCQB-UCLE